



CLASS II TYPE A

* Type A2 cabinets require all biologically contaminated ducts and plenums under negative pressure or surrounded by negative pressure ducts and plenums.



TITLE

**BIOLOGICAL SAFETY
CABINET
CLASS II, TYPE A**

FIGURE

VS-35-10

DATE

4-03